

BCMs and Charge For E12-06- 114:DVCS

Bishnu Karki
Ohio University



DVCS Collaboration Meeting

January 16, 2017

Outline:

- **Introduction:**
- **Unser / BCM Calibration:**
- **Charge Comparison:**
- **Conclusion:**

Introduction:

➤ Uncertainty in charge has significant impact on measured cross section.

➤ BCMs needs to be well calibrated for charge precision.

BCM Hardware:

➤ Two BCMs: UpStream (U) and DownStream(D)

➤ Multiple receivers for each BCM:

✓ **Analog:** X1 : U1 and D1


X3 : D3

X10 :D10

✓ **Digital:** Unew and Dnew

➤ Output from the BCM receivers are sent to Voltage to Frequency (V-F) converter then to Scaler.

➤ BCM Measures **relative current** so need to calibrated against Unser.





Unser / BCM Calibration:

Unser:

- Unser output is proportional to input current.
- Easy to calibrate.
- Uncertainty in Unser is $0.2 \mu\text{A}$.
- Unser pedestal drifts in unpredictable way over time scale of several minutes.

Calibration Procedure:


- ➔ First Calibrate Unser with known current source.
- ➔ Then use Unser current to calibrate BCMs.



Unser /BCM Calibration:

Unser Calibration:

- Unser is calibrated with known injected current in Hall A ($f_{\text{unser}} \rightarrow I_{\text{unser}}$).
- With precise knowledge of beam current from Unser, BCMs are calibrated ($f_{\text{BCM}} \rightarrow I_{\text{BCM}}$).



Unser / BCM Calibration:

Unser Gain Stability:

Unser gains Fall 2016

| Run | Gain X 10 ⁻⁶ (μ A/Hz) | Error X10 ⁻⁶ |
|-------|--|-------------------------|
| 23217 | 2506 | 5 |
| 23779 | 2504 | 5 |


Unser Gain Spring 2016

| Run | Gain X 10 ⁻⁶ (μ A/Hz) | Error X10 ⁻⁶ |
|-----------|--|-------------------------|
| Fall 2014 | 2755 | 7 |
| 21590 | 2754 | 6 |
| 12323 | 2753 | 6 |
| 22324 | 2753 | 6 |

Unser gains are stable within the error bar for particular run period.

Unser / BCM Calibration:


BCM Calibration:



➤ Using Unser current the BCMs are calibrated.

Unser / BCM Calibration:

Linearity Range:



| Device | U1 | D1 | D3 | D10 | Unew | Dnew |
|--------|-------|-------|------|------|-------------|-------------|
| I (μA) | 10-60 | 10-60 | 0-60 | 0-25 | 0-60 | 0-60 |

Stability Gains & Offsets:

Spring 2016:

| | |
|---------------|------------------------------|
| 12508 - 13015 | Gains from run 12514 & 12916 |
| 13100 - 13261 | Gains from run 13220 |
| 13279 - 13418 | Gains from run 13447 |

Spring 2016 Calibration Coefficient

Global Calibration from Spring 2016 by Zach

| BCM's | Gain (x 10 ⁻⁶ μA/Hz) | Offset (μA) |
|-------|---------------------------------|-------------|
| U1 | 351.17 ± 0.72 | 0.75 ± 0.06 |
| D1 | 319.28 ± 0.65 | 0.41 ± 0.06 |
| D3 | 93.09 ± 0.18 | 0.30 ± 0.05 |
| D10 | 32.14 ± 0.18 | 0.19 ± 0.06 |

Unew and Dnew (Run 12514 & 12916)









| | | |
|------|---------------|-------------|
| Dnew | 172.15 ± 0.66 | 0.19 ± 0.06 |
| Unew | 199.25 ± 0.76 | 0.20 ± 0.06 |

Unew and Dnew (Run 13220)

| | | |
|------|---------------|-------------|
| Dnew | 249.95 ± 1.41 | 0.10 ± 0.12 |
| Unew | 295.64 ± 1.66 | 0.20 ± 0.12 |

Unew and Dnew (Run 13447)

| | | |
|------|--------------|-------------|
| Dnew | 42.94 ± 0.25 | 0.04 ± 0.12 |
| Unew | 50.05 ± 0.29 | 0.05 ± 0.12 |



- The gains for new receivers changed multiple times.
- Analog BCMs: gains are stable so we can do global calibration.

Stability Gains & Offsets:

Fall 2016 & Fall 2014

Fall 2016 Calibration Coefficient

| October 15-2016 (Run 13852) High I (GMP) | | |
|--|--------------------------------------|-------------------|
| BCM's | Gain (x 10 ⁻⁶ μ A/Hz) | Offset (μ A) |
| U1 | 384.4 \pm 0.82 | 1.1 \pm 0.08 |
| D1 | 329.2 \pm 0.69 | 0.49 \pm 0.09 |
| D3 | 95.49 \pm 0.17 | 0.38 \pm 0.07 |
| D10 | 32.98 \pm 0.27 | 0.24 \pm 0.10 |
| Dnew | 215.54 \pm 0.38 | 0.13 \pm 0.07 |
| Unew | 254.6 \pm 0.45 | 0.15 \pm 0.07 |

| November 2-2016 (Run 14252) | | |
|-----------------------------|--------------------------------------|--------------------|
| BCM's | Gain (x 10 ⁻⁶ μ A/Hz) | Offset (μ A) |
| U1 | 383.1 \pm 3.43 | 1.28 \pm 0.16 |
| D1 | 326.3 \pm 2.93 | 0.81 \pm 0.16 |
| D3 | 97.7 \pm 0.60 | 0.13 \pm 0.09 |
| D10 | 34.36 \pm 0.32 | -0.12 \pm 0.11 |
| Dnew | 223.6 \pm 1.37 | -0.0003 \pm 0.09 |
| Unew | 258.1 \pm 1.6 | 0.003 \pm 0.09 |

Global Calibration (Nov 2 and Nov 26)

| BCM's | Gain (x 10 ⁻⁶ μ A/Hz) | Offset (μ A) |
|-------|--------------------------------------|-------------------|
| U1 | 384.84 \pm 1.86 | 1.10 \pm 0.11 |
| D1 | 328.77 \pm 1.59 | 0.62 \pm 0.11 |
| D3 | 97.05 \pm 0.32 | 0.19 \pm 0.06 |
| D10 | 33.72 \pm 0.22 | 0.03 \pm 0.08 |
| Dnew | 224.23 \pm 0.74 | -0.01 \pm 0.06 |
| Unew | 255.50 \pm 0.85 | 0.05 \pm 0.06 |

November 26-2016 (Run 14545)

| BCM's | Gain (x 10 ⁻⁶ μ A/Hz) | Offset (μ A) |
|-------|--------------------------------------|-------------------|
| U1 | 388.63 \pm 2.49 | 0.77 \pm 0.16 |
| D1 | 331.68 \pm 2.12 | 0.34 \pm 0.17 |
| D3 | 96.84 \pm 0.39 | 0.20 \pm 0.09 |
| D10 | 33.11 \pm 0.30 | 0.14 \pm 0.11 |
| Dnew | 224.37 \pm 0.91 | 0.0 \pm 0.09 |
| Unew | 254.89 \pm 1.04 | 0.03 \pm 0.09 |

Fall 2014 Calibration Coefficient


Dec 12 -2014 (Run 10505)

| BCM's | Gain (x 10 ⁻⁶ μ A/Hz) | Offset (μ A) |
|-------|--------------------------------------|-------------------|
| U1 | 515.9 \pm 9.99 | 0.56 \pm 0.29 |
| D1 | 454.53 \pm 8.8 | 0.55 \pm 0.22 |
| D3 | 127.4 \pm 1.6 | 0.39 \pm 0.17 |
| D10 | 45.65 \pm 0.57 | 0.031 \pm 0.17 |


- For Fall 2016, combined analysis from Nov 2 and Nov 26 can be used.
- Only one calibration run in Fall 2014 .

Charge Comparison:

- Charge (Q) = $\int I dt$




Fall 2014



Spring 2016

Charge Comparison:


Charge Comparison from Different BCMs



Fall 2016

- The charge from different BCMs agrees within 1%.
- New receivers are noisy and gains are not stable.
- U1 and D1 are not linear below 10 μA .
- D3 and D10 are linear within our range (5–20 μA) and gains are stable.

Slow Vs. Fast Scaler:



- **Charge from fast and Slow Scalers are in agreement for runs above 5 minutes.**

Gains and Offsets:

Fall 2016:

| Global Calibration (Nov 2 and Nov 26) | | |
|---------------------------------------|---------------------------------|--------------|
| BCM's | Gain (x 10 ⁻⁶ μA/Hz) | Offset (μA) |
| U1 | 384.84 ± 1.86 | 1.10 ± 0.11 |
| D1 | 328.77 ± 1.59 | 0.62 ± 0.11 |
| D3 | 97.05 ± 0.32 | 0.19 ± 0.06 |
| D10 | 33.72 ± 0.22 | 0.03 ± 0.08 |
| Dnew | 224.23 ± 0.74 | -0.01 ± 0.06 |
| Unew | 255.50 ± 0.85 | 0.05 ± 0.06 |

Fall 2014:

| Dec 12 -2014 (Run 10505) | | |
|--------------------------|---------------------------------|--------------|
| BCM's | Gain (x 10 ⁻⁶ μA/Hz) | Offset (μA) |
| U1 | 515.9 ± 9.99 | 0.56 ± 0.29 |
| D1 | 454.53 ± 8.8 | 0.55 ± 0.22 |
| D3 | 127.4 ± 1.6 | 0.39 ± 0.17 |
| D10 | 45.65 ± 0.57 | 0.031 ± 0.17 |


Spring 2016:

| Global Calibration from Spring 2016 by Zach | | |
|---|---------------------------------|-------------|
| BCM's | Gain (x 10 ⁻⁶ μA/Hz) | Offset (μA) |
| U1 | 351.17 ± 0.72 | 0.75 ± 0.06 |
| D1 | 319.28 ± 0.65 | 0.41 ± 0.06 |
| D3 | 93.09 ± 0.18 | 0.30 ± 0.05 |
| D10 | 32.14 ± 0.18 | 0.19 ± 0.06 |
| Unew and Dnew (Run 12514 & 12916) | | |
| Dnew | 172.15 ± 0.66 | 0.19 ± 0.06 |
| Unew | 199.25 ± 0.76 | 0.20 ± 0.06 |
| Unew and Dnew (Run 13220) | | |
| Dnew | 249.95 ± 1.41 | 0.10 ± 0.12 |
| Unew | 295.64 ± 1.66 | 0.20 ± 0.12 |
| Unew and Dnew (Run 13447) | | |
| Dnew | 42.94 ± 0.25 | 0.04 ± 0.12 |
| Unew | 50.05 ± 0.29 | 0.05 ± 0.12 |

| | |
|---------------|------------------------------|
| 12508 - 13015 | Gains from run 12514 & 12916 |
| 13100 - 13261 | Gains from run 13220 |
| 13279 - 13418 | Gains from run 13447 |

Conclusion:


- Unser gains are stable within a same run period.
- The gains for analog BCMs are stable within same run period so global calibration coefficient can be used for one run period.
- Different BCMs agrees with each other within 1%.
- D3 and D10 are linear in our current range and their gains are stable.
- Charge from slow and Fast Scaler are in agreement for run more than 5 minutes.





Thank you for your Attention



Unser Pedestal Variation, (No beam) Run 12813




BCMs Correlation:





Global Calibration

BCM vs UNSER, Combined analysis from Nov 2 and Nov 26 calibration




BCM vs UNSER, Combined analysis from Nov 2 and Nov 26 calibration







BCM vs UNSER, Combined analysis from Nov 2 and Nov 26 calibration




BCM vs UNSER, Combined analysis from Nov 2 and Nov 26 calibration



BCM Calibration Fall 2016




BCM vs UNSER, Combined analysis from Nov 2 and Nov 26 calibration



12

BCM vs UNSER, Combined analysis from Nov 2 and Nov 26 calibration



10

8

6

4

2

0





Gain Offsets Spring 2016

Spring 2016 Calibration Coefficient

| Feb-16 -2016 (Run 12514) | | | April-10-2016 (13220) | | |
|--------------------------|--------------------------|-------------------|-----------------------|--------------------------|-------------------|
| BCM's | Gain (x 10-6 μ A/Hz) | Offset (μ A) | BCM's | Gain (x 10-6 μ A/Hz) | Offset (μ A) |
| U1 | 354.3 \pm 3.5 | 0.57 \pm 0.22 | U1 | 352.9 \pm 2.1 | 0.68 \pm 0.16 |
| D1 | 322.3 \pm 3.2 | 0.55 \pm 0.22 | D1 | 320.2 \pm 1.88 | 0.35 \pm 0.16 |
| D3 | 93.9 \pm 0.57 | 0.19 \pm 0.12 | D3 | 93.38 \pm 0.65 | 0.29 \pm 0.13 |
| D10 | 32.2 \pm 0.33 | 0.14 \pm 0.14 | D10 | 32.09 \pm 0.49 | 0.19 \pm 0.19 |
| Dnew | 172.9 \pm 1.33 | 0.046 \pm 0.13 | Dnew | 249.0 \pm 1.6 | 0.15 \pm 0.12 |
| Unew | 200.7 \pm 1.21 | 0.034 \pm 0.12 | Unew | 294.3 \pm 1.4 | 0.25 \pm 0.12 |

| March-04-2016 (Run 12916) | | | April-21-2016 (13447) | | |
|---------------------------|--------------------------|-------------------|-----------------------|--------------------------|-------------------|
| BCM's | Gain (x 10-6 μ A/Hz) | Offset (μ A) | BCM's | Gain (x 10-6 μ A/Hz) | Offset (μ A) |
| U1 | 347.5 \pm 2.7 | 1.05 \pm 0.15 | U1 | 352.5 \pm 0.98 | 0.53 \pm 0.11 |
| D1 | 316.3 \pm 2.4 | 1.016 \pm 0.14 | D1 | 320.8 \pm 0.89 | 0.14 \pm 0.11 |
| D3 | 92.9 \pm 0.51 | 0.41 \pm 0.08 | D3 | 92.82 \pm 0.22 | 0.19 \pm 0.09 |
| D10 | 32.39 \pm 0.31 | 0.23 \pm 0.094 | D10 | 31.95 \pm 0.65 | 0.13 \pm 0.21 |
| Dnew | 172.7 \pm 0.94 | 0.23 \pm 0.08 | Dnew | 42.92 \pm 0.25 | 0.038 \pm 0.12 |
| Unew | 199.8 \pm 1.1 | 0.25 \pm 0.08 | Unew | 50.02 \pm 0.29 | 0.045 \pm 0.12 |


BCM Calibration Fall 2014




Charge comparison Fall 2016




Charge Comparison from Different BCMs



BCM calibration Spring 2016




BCM vs UNSER, Run: ALL (June 13, 2016)




BCM vs UNSER, Run: ALL

U1 Frequency verse


BCM vs UNSER, Run: ALL (June 13, 2016)




BCM vs UNSER, Run: ALL (June 13, 2016)




BCM vs UNSER, Run: ALL (June 13, 2016)



BCM vs UNSER, Run: ALL (June 13, 2016)



Residual U1



Charge Comparison Spring 2016

