



Preliminary results from Compton with FADC


 Rates up to 50 KHz with deadtim e of about 10%











DIS with pipelined DAQ

- SoLID plan to measure PVDIS asymmetries of order of a few 100 of PPM
- Counting DAQ using FADC
- Electronics procured: FADC and VTPs
- HMS and SHMS have Cerenkov and Calorimeters in FADC
- Proposal: add VTP to FADC crates measure electron and pion asymmetries from FADC with histograms on VTP
- Possible addition
 - Copy VDC signal and record tracks
- Parasitic of experiments

